

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
 Data File : VU051228.D
 Acq On : 07 Oct 2022 19:23
 Operator : JC/MD
 Sample : N4872-06DL 100X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXFX0DL

Quant Time: Oct 08 04:28:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 08 00:50:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	191143	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	208685	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	80129	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	73325	4.756	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.200%
7) Chloroethane-d5	1.909	69	68293	5.155	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.200%
11) 1,1-Dichloroethene-d2	2.562	65	28093	4.609	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.200%
20) 2-Butanone-d5	4.636	46	188887	55.643	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.280%
24) Chloroform-d	5.060	84	146651	5.102	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
26) 1,2-Dichloroethane-d4	5.700	65	77447	5.224	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.726	84	267749	4.664	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
36) 1,2-Dichloropropane-d6	6.690	67	95494	4.898	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.896	98	204569	3.983	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.600%
43) trans-1,3-Dichloroprop...	8.182	79	28681	4.478	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%
46) 2-Hexanone-d5	8.636	63	100640	46.617	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.240%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	74886	4.808	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	12.192	152	76387	5.485	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.600%
Target Compounds						
16) Methylene chloride	3.041	84	12225	0.594	ug/L	97
51) Chlorobenzene	9.446	112	45232	1.024	ug/L	98
64) 1,3-Dichlorobenzene	11.745	146	63916	2.546	ug/L	97
65) 1,4-Dichlorobenzene	11.835	146	229496	9.275	ug/L	98
67) 1,2-Dichlorobenzene	12.211	146	262072	10.738	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
Data File : VU051228.D
Acq On : 07 Oct 2022 19:23
Operator : JC/MD
Sample : N4872-06DL 100X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXFX0DL

Quant Time: Oct 08 04:28:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 08 00:50:53 2022
Response via : Initial Calibration

