

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
 Data File : VU051240.D
 Acq On : 08 Oct 2022 00:20
 Operator : JC/MD
 Sample : N4900-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXFT0

Quant Time: Oct 08 04:30:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 08 00:50:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	133444	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	148397	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	52119	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	55243	5.132	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 102.600%		
7) Chloroethane-d5	1.909	69	54438	5.886	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 117.800%		
11) 1,1-Dichloroethene-d2	2.565	65	21890	5.144	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 102.800%		
20) 2-Butanone-d5	4.620	46	151779	64.044	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 128.080%		
24) Chloroform-d	5.060	84	121425	6.051	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 121.000%		
26) 1,2-Dichloroethane-d4	5.700	65	69767	6.741	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 134.800%#		
32) Benzene-d6	5.726	84	214154	5.246	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 105.000%		
36) 1,2-Dichloropropane-d6	6.690	67	79514	5.735	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 114.800%		
41) Toluene-d8	7.896	98	161273	4.416	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 88.400%		
43) trans-1,3-Dichloroprop...	8.179	79	22719	4.988	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 99.800%		
46) 2-Hexanone-d5	8.632	63	78038	50.833	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 101.660%		
56) 1,1,2,2-Tetrachloroeth...	10.754	84	67665	6.110	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 122.200%#		
66) 1,2-Dichlorobenzene-d4	12.192	152	57002	6.292	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 125.800%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
Data File : VU051240.D
Acq On : 08 Oct 2022 00:20
Operator : JC/MD
Sample : N4900-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXFT0

Quant Time: Oct 08 04:30:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 08 00:50:53 2022
Response via : Initial Calibration

