

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
 Data File : VU051247.D
 Acq On : 08 Oct 2022 03:38
 Operator : JC/MD
 Sample : N4900-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXFT4

Quant Time: Oct 08 23:47:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 08 05:40:36 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.250	114	140196	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.417	117	158324	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.812	152	54055	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.597	65	57029	5.043	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.800%	
7) Chloroethane-d5		1.909	69	57006	5.867	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.400%	
11) 1,1-Dichloroethene-d2		2.562	65	22652	5.067	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.400%	
20) 2-Butanone-d5		4.623	46	138172	55.495	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.980%	
24) Chloroform-d		5.063	84	127505	6.048	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	121.000%	
26) 1,2-Dichloroethane-d4		5.700	65	68706	6.319	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	126.400%	
32) Benzene-d6		5.726	84	227471	5.223	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.400%	
36) 1,2-Dichloropropane-d6		6.690	67	82258	5.561	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.200%	
41) Toluene-d8		7.896	98	168744	4.330	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.600%	
43) trans-1,3-Dichloroprop...		8.179	79	22865	4.705	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.200%	
46) 2-Hexanone-d5		8.632	63	67935	41.478	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.960%	
56) 1,1,2,2-Tetrachloroeth...		10.755	84	65436	5.538	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.800%	
66) 1,2-Dichlorobenzene-d4		12.192	152	56227	5.985	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.600%	
Target Compounds							
13) Acetone		2.629	43	5218	2.794	ug/L	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
Data File : VU051247.D
Acq On : 08 Oct 2022 03:38
Operator : JC/MD
Sample : N4900-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXFT4

Quant Time: Oct 08 23:47:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 08 05:40:36 2022
Response via : Initial Calibration

