

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
 Data File : VU051259.D
 Acq On : 08 Oct 2022 08:35
 Operator : JC/MD
 Sample : N4931-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5H06

Quant Time: Oct 08 23:48:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 08 05:40:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	162214	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	181979	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	64792	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	0.000	65	0d	0.000	ug/L	
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.000%#
7) Chloroethane-d5	1.909	69	55951	4.977	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.600%
11) 1,1-Dichloroethene-d2	0.000	65	0d	0.000	ug/L	
Spiked Amount	5.000	Range	60 - 125	Recovery	=	0.000%#
20) 2-Butanone-d5	4.630	46	152127	52.806	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.620%
24) Chloroform-d	5.060	84	121899	4.997	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.000%
26) 1,2-Dichloroethane-d4	5.703	65	65960	5.243	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.800%
32) Benzene-d6	5.726	84	202773	4.050	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.000%
36) 1,2-Dichloropropane-d6	6.691	67	77791	4.575	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.600%
41) Toluene-d8	7.900	98	1251	0.028	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.600%#
43) trans-1,3-Dichloroprop...	0.000	79	0	0.000	ug/L	
Spiked Amount	5.000	Range	55 - 130	Recovery	=	0.000%#
46) 2-Hexanone-d5	8.633	63	70586	37.494	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.980%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	65726	4.839	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.800%
66) 1,2-Dichlorobenzene-d4	12.192	152	56607	5.027	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%
Target Compounds						
					Qvalue	
38) Bromodichloromethane	7.105	83	13446	0.677	ug/L	94
49) Dibromochloromethane	8.809	129	19915	1.517	ug/L	96
59) Bromoform	10.288	173	11589	2.150	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
Data File : VU051259.D
Acq On : 08 Oct 2022 08:35
Operator : JC/MD
Sample : N4931-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5H06

Quant Time: Oct 08 23:48:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 08 05:40:36 2022
Response via : Initial Calibration

