

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
 Data File : VU051273.D
 Acq On : 08 Oct 2022 15:40
 Operator : JC/MD
 Sample : VIBLK112
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK112

Quant Time: Oct 08 23:50:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 08 05:40:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	171472	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	190178	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	69794	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	62597	4.526	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.600%
7) Chloroethane-d5	1.912	69	60754	5.112	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.200%
11) 1,1-Dichloroethene-d2	2.565	65	24999	4.572	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.400%
20) 2-Butanone-d5	4.623	46	161793	53.129	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.260%
24) Chloroform-d	5.063	84	136201	5.282	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.600%
26) 1,2-Dichloroethane-d4	5.703	65	72988	5.488	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.800%
32) Benzene-d6	5.726	84	242313	4.632	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
36) 1,2-Dichloropropane-d6	6.690	67	87061	4.900	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.896	98	180109	3.848	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.000%
43) trans-1,3-Dichloroprop...	8.179	79	22694	3.888	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.800%
46) 2-Hexanone-d5	8.632	63	83381	42.381	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.760%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	71204	5.017	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.400%
66) 1,2-Dichlorobenzene-d4	12.192	152	62883	5.184	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%
Target Compounds						
					Qvalue	
13) Acetone	2.626	43	6945	3.041	ug/L	96
16) Methylene chloride	3.044	84	9762	0.529	ug/L	96
51) Chlorobenzene	9.446	112	4611	0.115	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
Data File : VU051273.D
Acq On : 08 Oct 2022 15:40
Operator : JC/MD
Sample : VIBLK112
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 63 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK112

Quant Time: Oct 08 23:50:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 08 05:40:36 2022
Response via : Initial Calibration

