

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
 Data File : VU051288.D
 Acq On : 08 Oct 2022 21:49
 Operator : JC/MD
 Sample : N4900-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXFT0

Quant Time: Oct 08 23:52:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 08 05:40:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	167924	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	192781	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	72613	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	58333	4.306	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.200%	
7) Chloroethane-d5	1.912	69	56096	4.820	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.400%	
11) 1,1-Dichloroethene-d2	2.565	65	24239	4.527	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.600%	
20) 2-Butanone-d5	4.620	46	145550	48.805	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	97.620%	
24) Chloroform-d	5.063	84	130572	5.170	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.400%	
26) 1,2-Dichloroethane-d4	5.703	65	70081	5.381	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.600%	
32) Benzene-d6	5.726	84	226760	4.276	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.600%	
36) 1,2-Dichloropropane-d6	6.690	67	83893	4.658	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.200%	
41) Toluene-d8	7.896	98	169784	3.578	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.600%	
43) trans-1,3-Dichloroprop...	8.179	79	23322	3.941	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.800%	
46) 2-Hexanone-d5	8.632	63	73859	37.034	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	74.060%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	68814	4.783	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.600%	
66) 1,2-Dichlorobenzene-d4	12.192	152	63249	5.011	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
Data File : VU051288.D
Acq On : 08 Oct 2022 21:49
Operator : JC/MD
Sample : N4900-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXFT0

Quant Time: Oct 08 23:52:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 08 05:40:36 2022
Response via : Initial Calibration

