

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
 Data File : VU051239.D
 Acq On : 07 Oct 2022 23:56
 Operator : JC/MD
 Sample : N4900-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXFS9

Quant Time: Oct 08 04:30:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 08 00:50:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	131241	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	149697	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	50978	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	57071	5.391	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	107.800%
7) Chloroethane-d5	1.909	69	55252	6.074	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.400%
11) 1,1-Dichloroethene-d2	2.562	65	20802	4.971	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.400%
20) 2-Butanone-d5	4.620	46	156281	67.051	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	134.100%#
24) Chloroform-d	5.060	84	125067	6.337	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	126.800%#
26) 1,2-Dichloroethane-d4	5.700	65	68279	6.708	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	134.200%#
32) Benzene-d6	5.726	84	216859	5.266	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.400%
36) 1,2-Dichloropropane-d6	6.690	67	82197	5.877	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.600%
41) Toluene-d8	7.896	98	163277	4.432	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.600%
43) trans-1,3-Dichloroprop...	8.179	79	22643	4.928	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.600%
46) 2-Hexanone-d5	8.632	63	79945	51.623	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.240%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	71240	6.377	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	127.600%#
66) 1,2-Dichlorobenzene-d4	12.192	152	56382	6.363	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	127.200%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
Data File : VU051239.D
Acq On : 07 Oct 2022 23:56
Operator : JC/MD
Sample : N4900-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXFS9

Quant Time: Oct 08 04:30:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 08 00:50:53 2022
Response via : Initial Calibration

