

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101122\
 Data File : VU051377.D
 Acq On : 11 Oct 2022 17:17
 Operator : JC/MD
 Sample : N5041-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 11 23:39:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 11 23:36:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	176764	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	205397	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	74454	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	41105	2.883	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	57.600%
7) Chloroethane-d5	1.909	69	44964	3.670	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.400%
11) 1,1-Dichloroethene-d2	2.562	65	19667	3.489	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.800%
20) 2-Butanone-d5	4.639	46	194900	62.085	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.160%
24) Chloroform-d	5.060	84	133188	5.010	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.700	65	74644	5.445	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.800%
32) Benzene-d6	5.726	84	218844	3.873	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.400%
36) 1,2-Dichloropropane-d6	6.690	67	87765	4.573	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.400%
41) Toluene-d8	7.896	98	157271	3.111	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	62.200%#
43) trans-1,3-Dichloroprop...	8.179	79	24883	3.947	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.000%
46) 2-Hexanone-d5	8.636	63	104501	49.180	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.360%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	81459	5.314	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.200%
66) 1,2-Dichlorobenzene-d4	12.192	152	62069	4.796	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%
Target Compounds						
16) Methylene chloride	3.041	84	12211	0.642	ug/L	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101122\
Data File : VU051377.D
Acq On : 11 Oct 2022 17:17
Operator : JC/MD
Sample : N5041-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 11 23:39:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 11 23:36:07 2022
Response via : Initial Calibration

