

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101122\
 Data File : VU051368.D
 Acq On : 11 Oct 2022 11:00
 Operator : JC/MD
 Sample : VU1011WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK112

Quant Time: Oct 11 23:37:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 11 23:36:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	202945	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	229209	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	86501	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	75458	4.609	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.200%
7) Chloroethane-d5	1.909	69	70076	4.982	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.600%
11) 1,1-Dichloroethene-d2	2.562	65	29518	4.561	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.200%
20) 2-Butanone-d5	4.626	46	188445	52.284	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.560%
24) Chloroform-d	5.060	84	153467	5.028	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
26) 1,2-Dichloroethane-d4	5.700	65	80727	5.129	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.600%
32) Benzene-d6	5.726	84	267670	4.245	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
36) 1,2-Dichloropropane-d6	6.690	67	99219	4.633	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.600%
41) Toluene-d8	7.896	98	196170	3.477	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.600%#
43) trans-1,3-Dichloroprop...	8.179	79	29633	4.212	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.200%
46) 2-Hexanone-d5	8.632	63	103747	43.753	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.500%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	81977	4.792	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	12.192	152	72626	4.831	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101122\
Data File : VU051368.D
Acq On : 11 Oct 2022 11:00
Operator : JC/MD
Sample : VU1011WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK112

Quant Time: Oct 11 23:37:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 11 23:36:07 2022
Response via : Initial Calibration

