



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101122\
 Data File : VU051366.D
 Acq On : 11 Oct 2022 09:15
 Operator : JC/MD
 Sample : BFB154
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB154

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Thu Oct 13 01:37:29 2022



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	49283	PASS
75	95	30	80	49.6	132864	PASS
95	95	100	100	100.0	267947	PASS
96	95	5	9	6.5	17344	PASS
173	174	0.00	2	1.5	3038	PASS
174	95	50	120	73.3	196501	PASS
175	174	5	9	7.6	15030	PASS
176	174	95	101	97.0	190613	PASS
177	176	5	9	6.4	12108	PASS

m/z	Abundance
39.80	375.0
43.90	456.0

Instrument :
MSVOA_U
ClientSampleId :
BFB154

m/z	Abundance
36.00	1907.0
37.00	11614.0
38.00	10084.0
39.00	5044.0
40.00	680.0
43.00	167.0
44.00	1920.0
45.00	2189.0
45.90	565.0
46.90	2898.0
48.00	1558.0
49.00	9845.0
50.00	47568.0
51.00	14538.0
52.00	692.0
53.00	395.0
54.00	262.0
54.90	772.0
56.00	3873.0
57.00	6760.0
58.00	417.0
60.00	2527.0
61.00	12432.0
62.00	13502.0
63.00	12150.0
64.00	2742.0
65.00	2503.0
66.00	395.0
67.00	612.0
68.00	25616.0
69.00	26792.0
70.00	1992.0
71.00	105.0
72.00	1290.0
73.00	11584.0
74.00	43144.0
75.00	128640.0
76.00	11774.0
77.00	1955.0
77.90	920.0
78.90	11740.0
79.90	2964.0
80.90	10827.0
81.90	2245.0
83.00	318.0
85.80	435.0
86.90	11489.0
88.00	10945.0
90.90	970.0
92.00	7254.0
93.00	11162.0
94.00	30032.0
95.00	254720.0
96.00	16664.0
96.90	480.0
102.90	102.0
103.90	1313.0
104.80	404.0
105.90	1067.0
106.90	192.0
107.90	295.0
109.60	122.0
110.00	104.0
110.80	165.0
113.00	142.0
114.90	267.0
115.90	833.0
116.90	2293.0
117.90	858.0
118.90	1816.0
123.80	150.0
124.80	133.0
125.90	114.0
127.90	886.0
128.90	456.0
129.90	841.0
130.90	303.0
133.90	177.0
134.90	499.0
136.90	491.0
139.90	148.0
140.90	2165.0
141.90	249.0
143.00	4734.0
144.00	366.0
144.90	2810.0
145.90	556.0
146.80	196.0
147.90	499.0
148.90	123.0
149.90	190.0
151.80	117.0

Instrument :
MSVOA_U
ClientSampleId :
BFB154

152.90	129.0
154.90	472.0
155.90	122.0
156.80	450.0
159.00	160.0
160.90	291.0
170.90	1011.0
171.90	3106.0
172.90	2892.0
174.00	183232.0
175.00	14177.0
176.00	177920.0
177.00	11554.0
177.90	317.0

Instrument :
MSVOA_U
ClientSampleId :
BFB154

m/z	Abundance
36.00	1818.0
37.00	12217.0
38.00	10831.0
39.00	5157.0
40.00	637.0
43.00	168.0
43.90	2034.0
44.90	2337.0
45.90	553.0
46.90	2906.0
48.00	1520.0
49.00	10529.0
50.00	50808.0
51.00	15723.0
52.00	677.0
53.00	430.0
54.00	261.0
54.90	837.0
56.00	4316.0
57.00	6909.0
58.00	329.0
60.00	2795.0
61.00	13590.0
62.00	14543.0
63.00	12914.0
64.00	2750.0
65.00	2761.0
66.00	342.0
66.90	638.0
68.00	27424.0
69.00	28472.0
70.00	2106.0
72.00	1569.0
73.00	12638.0
74.00	46416.0
75.00	136832.0
76.00	12534.0
77.00	1986.0
77.90	916.0
78.90	12530.0
79.90	3006.0
80.90	10943.0
81.90	2455.0
83.00	348.0
85.90	495.0
86.90	12230.0
88.00	12249.0
90.90	1087.0
92.00	7778.0
93.00	12021.0
94.00	32296.0
95.00	277056.0
96.00	17832.0
97.00	523.0
103.90	1294.0
104.80	469.0
105.90	1098.0
106.90	222.0
107.90	324.0
110.00	168.0
110.80	193.0
112.00	115.0
113.00	146.0
114.90	263.0
115.90	885.0
116.90	2536.0
118.00	944.0
118.90	1982.0
123.70	151.0
124.80	112.0
125.80	150.0
127.90	956.0
128.90	488.0
129.90	1088.0
130.90	315.0
133.90	165.0
134.90	509.0
136.90	496.0
139.80	170.0
140.90	2592.0
142.00	281.0
142.90	4809.0
144.00	427.0
144.90	2603.0
145.90	646.0
146.80	230.0
147.90	539.0
148.80	115.0
149.90	200.0
151.80	116.0
152.90	110.0
153.70	126.0

Instrument :
MSVOA_U
ClientSampleId :
BFB154

154.90	589.0
156.90	494.0
159.00	245.0
160.90	326.0
170.80	967.0
171.90	3489.0
172.90	3221.0
174.00	202688.0
175.00	15684.0
176.00	196864.0
177.00	12502.0
177.90	348.0

Instrument :
MSVOA_U
ClientSampleId :
BFB154

m/z	Abundance
36.00	1745.0
37.00	11805.0
38.00	10497.0
39.00	4570.0
39.90	511.0
40.10	513.0
43.00	169.0
43.90	1911.0
44.90	2231.0
45.90	461.0
46.90	2842.0
47.90	1518.0
49.00	10492.0
50.00	49472.0
51.00	15366.0
52.00	557.0
53.00	456.0
54.00	226.0
54.90	778.0
56.00	4153.0
57.00	6520.0
57.90	225.0
60.00	2713.0
61.00	13545.0
62.00	14118.0
63.00	12186.0
64.00	2541.0
65.00	2714.0
66.00	210.0
66.90	657.0
68.00	26288.0
69.00	27392.0
70.00	2195.0
71.90	1593.0
73.00	12457.0
74.00	45720.0
75.00	133120.0
76.00	12112.0
77.00	1777.0
77.90	856.0
78.90	11789.0
79.90	2900.0
80.90	10536.0
81.90	2494.0
83.00	342.0
85.90	538.0
87.00	12003.0
88.00	11815.0
90.90	994.0
92.00	7756.0
93.00	11943.0
94.00	31432.0
95.00	272064.0
96.00	17536.0
97.00	516.0
103.90	1153.0
104.90	460.0
105.90	1004.0
106.80	292.0
107.90	334.0
109.90	189.0
110.90	228.0
112.00	141.0
113.00	117.0
114.80	266.0
115.80	904.0
116.90	2469.0
117.90	965.0
118.90	2017.0
122.90	127.0
123.70	152.0
124.90	111.0
125.80	151.0
127.90	921.0
128.90	493.0
129.90	1193.0
130.90	275.0
133.90	116.0
134.90	424.0
136.80	455.0
139.70	174.0
140.90	2598.0
141.90	263.0
142.90	4596.0
144.00	378.0
144.90	2257.0
146.00	671.0
146.80	188.0
147.90	497.0
148.80	138.0
149.90	187.0
152.90	130.0

Instrument :
MSVOA_U
ClientSampleId :
BFB154

153.70	119.0
154.90	624.0
156.90	474.0
158.90	255.0
160.90	334.0
170.80	892.0
171.90	3490.0
172.90	3002.0
174.00	203584.0
175.00	15230.0
176.00	197056.0
177.00	12268.0
177.90	349.0

Instrument :
MSVOA_U
ClientSampleId :
BFB154