

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110222\
 Data File : VU051685.D
 Acq On : 02 Nov 2022 12:22
 Operator : JC/MD
 Sample : N5353-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBWR2

Quant Time: Nov 03 01:47:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 03 01:45:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	170858	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	182113	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	72028	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	70379	5.469	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 109.400%		
7) Chloroethane-d5	1.909	69	63431	4.726	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 94.600%		
11) 1,1-Dichloroethene-d2	2.565	65	28325	6.065	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 121.400%		
20) 2-Butanone-d5	4.620	46	164544	35.307	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 70.620%		
24) Chloroform-d	5.060	84	135301	4.493	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 89.800%		
26) 1,2-Dichloroethane-d4	5.700	65	71649	4.405	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 88.000%		
32) Benzene-d6	5.726	84	265367	4.618	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 92.400%		
36) 1,2-Dichloropropane-d6	6.690	67	89965	4.477	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 89.600%		
41) Toluene-d8	7.896	98	204289	4.520	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 90.400%		
43) trans-1,3-Dichloroprop...	8.179	79	25787	4.197	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 84.000%		
46) 2-Hexanone-d5	8.632	63	120075	32.462	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 64.920%		
56) 1,1,2,2-Tetrachloroeth...	10.754	84	77371	4.201	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 84.000%		
66) 1,2-Dichlorobenzene-d4	12.192	152	73300	5.545	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 110.800%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110222\
Data File : VU051685.D
Acq On : 02 Nov 2022 12:22
Operator : JC/MD
Sample : N5353-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBWR2

Quant Time: Nov 03 01:47:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 03 01:45:18 2022
Response via : Initial Calibration

