

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110222\
 Data File : VU051704.D
 Acq On : 02 Nov 2022 21:08
 Operator : JC/MD
 Sample : VU1102WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK129

Quant Time: Nov 03 03:06:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 03 03:06:09 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	163557	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	174759	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	73704	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	65273	5.299	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 106.000%		
7) Chloroethane-d5	1.912	69	61369	4.777	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 95.600%		
11) 1,1-Dichloroethene-d2	2.565	65	26740	5.981	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 119.600%		
20) 2-Butanone-d5	4.639	46	165942	37.197	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 74.400%		
24) Chloroform-d	5.063	84	133655	4.637	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 92.800%		
26) 1,2-Dichloroethane-d4	5.700	65	67875	4.359	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 87.200%		
32) Benzene-d6	5.726	84	265432	4.813	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 96.200%		
36) 1,2-Dichloropropane-d6	6.690	67	87937	4.560	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 91.200%		
41) Toluene-d8	7.896	98	202547	4.670	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 93.400%		
43) trans-1,3-Dichloroprop...	8.179	79	25551	4.333	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 86.600%		
46) 2-Hexanone-d5	8.632	63	120165	33.854	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 67.700%		
56) 1,1,2,2-Tetrachloroeth...	10.755	84	72642	4.111	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 82.200%		
66) 1,2-Dichlorobenzene-d4	12.192	152	72679	5.373	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 107.400%		
Target Compounds						
16) Methylene chloride	3.044	84	7518	0.323	ug/L	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110222\
Data File : VU051704.D
Acq On : 02 Nov 2022 21:08
Operator : JC/MD
Sample : VU1102WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK129

Quant Time: Nov 03 03:06:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 03 03:06:09 2022
Response via : Initial Calibration

