

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110222\
 Data File : VU051713.D
 Acq On : 03 Nov 2022 01:00
 Operator : JC/MD
 Sample : N5301-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBWP4

Quant Time: Nov 03 04:29:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 03 03:06:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	155450	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	177369	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	78173	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	62219	5.314	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 106.200%		
7) Chloroethane-d5	1.912	69	58714	4.809	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 96.200%		
11) 1,1-Dichloroethene-d2	2.565	65	23935	5.633	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 112.600%		
20) 2-Butanone-d5	4.620	46	187233	44.158	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 88.320%		
24) Chloroform-d	5.060	84	139199	5.081	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 101.600%		
26) 1,2-Dichloroethane-d4	5.700	65	71022	4.799	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 96.000%		
32) Benzene-d6	5.726	84	253445	4.528	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 90.600%		
36) 1,2-Dichloropropane-d6	6.690	67	89018	4.548	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 91.000%		
41) Toluene-d8	7.896	98	188908	4.292	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 85.800%		
43) trans-1,3-Dichloroprop...	8.179	79	26526	4.432	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 88.600%		
46) 2-Hexanone-d5	8.632	63	139352	38.681	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 77.360%		
56) 1,1,2,2-Tetrachloroeth...	10.755	84	85030	4.741	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 94.800%		
66) 1,2-Dichlorobenzene-d4	12.192	152	77569	5.406	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 108.200%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110222\
Data File : VU051713.D
Acq On : 03 Nov 2022 01:00
Operator : JC/MD
Sample : N5301-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBWP4

Quant Time: Nov 03 04:29:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 03 03:06:09 2022
Response via : Initial Calibration

