

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112222\
 Data File : VU052057.D
 Acq On : 22 Nov 2022 22:30
 Operator : JC/MD
 Sample : N5684-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH1M8

Quant Time: Nov 23 00:50:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 19 21:39:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	217695	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	229453	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	87828	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	104630	4.941	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	98.800%	
7) Chloroethane-d5	1.912	69	85685	5.210	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.200%	
11) 1,1-Dichloroethene-d2	2.565	65	37615	4.956	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	99.200%	
20) 2-Butanone-d5	4.626	46	193737	50.475	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.940%	
24) Chloroform-d	5.060	84	163096	5.242	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.800%	
26) 1,2-Dichloroethane-d4	5.700	65	83902	5.339	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.800%	
32) Benzene-d6	5.726	84	334664	4.913	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.200%	
36) 1,2-Dichloropropane-d6	6.687	67	107225	5.086	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.800%	
41) Toluene-d8	7.896	98	262776	4.264	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.200%	
43) trans-1,3-Dichloroprop...	8.179	79	27262	3.549	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	71.000%	
46) 2-Hexanone-d5	8.632	63	137630	42.274	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	84.540%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	87936	4.811	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.200%	
66) 1,2-Dichlorobenzene-d4	12.192	152	94516	5.621	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	112.400%	
Target Compounds						Qvalue
3) Chloromethane	1.523	50	6378	0.259	ug/L	95
13) Acetone	2.633	43	45521	17.465	ug/L	99
34) Trichloroethene	6.546	95	3826	0.205	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112222\
Data File : VU052057.D
Acq On : 22 Nov 2022 22:30
Operator : JC/MD
Sample : N5684-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH1M8

Quant Time: Nov 23 00:50:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 19 21:39:41 2022
Response via : Initial Calibration

