

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112222\
 Data File : VU052074.D
 Acq On : 23 Nov 2022 07:07
 Operator : JC/MD
 Sample : N5693-10DL 400X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBN91DL

Quant Time: Nov 23 08:26:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 19 21:39:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	186732	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	189793	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	68061	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	92029	5.067	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	101.400%	
7) Chloroethane-d5	1.916	69	79539	5.638	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	112.800%	
11) 1,1-Dichloroethene-d2	2.565	65	33796	5.191	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	103.800%	
20) 2-Butanone-d5	4.633	46	215589	65.482	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	130.960%#	
24) Chloroform-d	5.064	84	157716	5.910	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	118.200%	
26) 1,2-Dichloroethane-d4	5.700	65	83208	6.173	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	123.400%	
32) Benzene-d6	5.726	84	315843	5.605	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	112.200%	
36) 1,2-Dichloropropane-d6	6.690	67	106227	6.092	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	121.800%	
41) Toluene-d8	7.896	98	236216	4.633	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.600%	
43) trans-1,3-Dichloroprop...	8.179	79	32468	5.109	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	102.200%	
46) 2-Hexanone-d5	8.633	63	157772	58.587	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	117.180%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	89882	5.945	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	119.000%	
66) 1,2-Dichlorobenzene-d4	12.192	152	80118	6.148	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	123.000%#	
Target Compounds						Qvalue
16) Methylene chloride	3.048	84	13494	0.736	ug/L	98
22) cis-1,2-Dichloroethene	4.662	96	20294	1.332	ug/L	93
34) Trichloroethene	6.539	95	92705	6.017	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112222\
Data File : VU052074.D
Acq On : 23 Nov 2022 07:07
Operator : JC/MD
Sample : N5693-10DL 400X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBN91DL

Quant Time: Nov 23 08:26:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 19 21:39:41 2022
Response via : Initial Calibration

