




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112222\
 Data File : VU052070.D
 Acq On : 23 Nov 2022 05:18
 Operator : JC/MD
 Sample : N5693-06DL 10X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBN87DL

Manual IntegrationsAPPROVED

Quant Time: Nov 23 08:26:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 19 21:39:41 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 11/23/2022
 Supervised By :Mahesh Dadoda 11/25/2022



TIC: VU052070.D\data.ms

(13) Acetone (T)

2.662min (+ 0.042) 11.94 ug/L

response 29809

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	38.00	27.35
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112222\
 Data File : VU052070.D
 Acq On : 23 Nov 2022 05:18
 Operator : JC/MD
 Sample : N5693-06DL 10X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBN87DL

Manual IntegrationsAPPROVED

Quant Time: Nov 23 08:26:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 19 21:39:41 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 11/23/2022
 Supervised By :Mahesh Dadoda 11/25/2022



TIC: VU052070.D\data.ms

(13) Acetone (T)

2.662min (+ 0.042) 12.18 ug/L m

response 30411

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	38.00	26.81
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112222\
 Data File : VU052070.D
 Acq On : 23 Nov 2022 05:18
 Operator : JC/MD
 Sample : N5693-06DL 10X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBN87DL

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 11/23/2022
 Supervised By :Mahesh Dadoda 11/25/2022

Quant Time: Nov 23 08:26:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 19 21:39:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	208493	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	217016	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	78977	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	95037	4.686	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.800%
7) Chloroethane-d5	1.912	69	82212	5.219	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.400%
11) 1,1-Dichloroethene-d2	2.565	65	36055	4.960	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.200%
20) 2-Butanone-d5	4.639	46	233576	63.540	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.080%
24) Chloroform-d	5.060	84	171472	5.754	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.000%
26) 1,2-Dichloroethane-d4	5.700	65	90147	5.989	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.800%
32) Benzene-d6	5.726	84	349564	5.426	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.600%
36) 1,2-Dichloropropane-d6	6.690	67	113174	5.676	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.600%
41) Toluene-d8	7.896	98	272241	4.670	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
43) trans-1,3-Dichloroprop...	8.179	79	36229	4.986	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.800%
46) 2-Hexanone-d5	8.636	63	167347	54.347	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.700%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	96174	5.563	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.200%
66) 1,2-Dichlorobenzene-d4	12.192	152	91852	6.075	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	121.400%#
Target Compounds					Qvalue	
13) Acetone	2.662	43	30411m	12.183	ug/L	
16) Methylene chloride	3.047	84	15210	0.743	ug/L	97
22) cis-1,2-Dichloroethene	4.665	96	24103	1.417	ug/L	92
34) Trichloroethene	6.539	95	153270	8.701	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112222\
Data File : VU052070.D
Acq On : 23 Nov 2022 05:18
Operator : JC/MD
Sample : N5693-06DL 10X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBN87DL

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 11/23/2022
Supervised By :Mahesh Dadoda 11/25/2022

Quant Time: Nov 23 08:26:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 19 21:39:41 2022
Response via : Initial Calibration

