


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010220\
 Data File : VU036387.D
 Acq On : 02 Jan 2020 10:38
 Operator : JC/MD
 Sample : BFB28
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB28

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122819WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Jan 03 17:51:02 2020



AutoFind: Scans 2897, 2898, 2899; Background Corrected with Scan 2882

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	30109	PASS
75	95	30	80	53.8	83549	PASS
95	95	100	100	100.0	155242	PASS
96	95	5	9	6.4	9876	PASS
173	174	0.00	2	0.6	633	PASS
174	95	50	120	73.9	114794	PASS
175	174	5	9	7.2	8272	PASS
176	174	95	101	96.7	110992	PASS
177	176	5	9	6.5	7229	PASS

m/z	Abundance
38.90	205.0
40.00	267.0
40.90	181.0
41.80	261.0
42.30	106.0
43.00	224.0
43.80	1166.0
51.00	106.0
54.90	297.0
55.80	115.0
56.90	108.0
66.90	175.0
70.00	113.0
72.10	107.0
81.20	100.0
82.80	151.0
91.10	116.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

m/z	Abundance
35.90	1294.0
36.90	8118.0
37.90	7516.0
38.90	3299.0
39.90	640.0
41.00	455.0
41.80	127.0
42.70	161.0
43.10	191.0
43.90	2615.0
44.90	1577.0
45.90	342.0
46.90	1449.0
47.90	1048.0
49.00	6502.0
49.90	31048.0
50.90	10507.0
51.90	291.0
54.80	367.0
55.00	374.0
55.90	3090.0
57.00	5163.0
57.90	363.0
59.90	1765.0
61.00	8576.0
61.90	8594.0
62.90	7124.0
64.00	584.0
64.90	187.0
66.90	704.0
68.00	16808.0
68.90	18904.0
69.90	1653.0
71.90	1080.0
73.00	7382.0
74.00	30896.0
74.90	86248.0
76.00	6859.0
77.00	937.0
77.80	560.0
78.90	7978.0
79.90	2016.0
80.90	7430.0
81.90	1590.0
83.00	311.0
87.00	6330.0
87.90	5954.0
90.70	790.0
92.00	5420.0
93.00	7101.0
94.00	19648.0
95.00	160128.0
96.00	10134.0
96.90	317.0
103.80	1006.0
105.00	205.0
105.90	935.0
106.60	185.0
107.10	123.0
109.70	138.0
110.80	408.0
111.60	138.0
112.70	270.0
114.70	290.0
115.70	703.0
116.90	1268.0
117.80	669.0
118.80	1013.0
127.70	611.0
128.90	170.0
129.90	606.0
130.90	208.0
134.90	471.0
136.90	462.0
140.80	2193.0
141.90	346.0
142.80	2067.0
143.90	124.0
144.80	116.0
145.80	238.0
146.90	152.0
148.00	356.0
148.80	184.0
151.80	133.0
153.00	153.0
154.90	392.0
156.80	271.0
160.70	235.0
169.20	146.0
170.70	263.0
172.10	411.0
172.70	892.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

173.00	913.0
173.90	117928.0
174.90	8498.0
175.90	114048.0
176.90	7322.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

m/z	Abundance
35.90	1086.0
36.90	7882.0
37.90	7166.0
38.90	3228.0
39.90	560.0
41.00	397.0
42.70	214.0
43.90	2719.0
44.90	1566.0
46.00	276.0
46.90	1349.0
48.00	1053.0
49.00	6490.0
49.90	29712.0
50.90	10129.0
52.00	325.0
55.00	490.0
56.00	3179.0
56.90	5155.0
57.80	320.0
59.90	1722.0
61.00	8397.0
61.90	8019.0
62.90	7093.0
64.00	482.0
64.90	169.0
66.90	622.0
67.90	16704.0
68.90	18240.0
69.90	1629.0
70.70	142.0
72.00	967.0
72.90	6922.0
74.00	30848.0
75.00	83992.0
76.00	6831.0
77.00	993.0
77.90	546.0
78.90	7783.0
79.80	1992.0
80.90	7470.0
81.80	1635.0
83.00	324.0
86.90	5966.0
87.80	6309.0
90.70	717.0
91.90	5229.0
93.00	7024.0
94.00	19504.0
95.00	156992.0
96.00	9929.0
96.90	295.0
103.80	1083.0
104.90	246.0
105.90	940.0
106.70	208.0
107.10	224.0
109.80	142.0
110.80	290.0
111.60	140.0
112.70	211.0
114.70	252.0
115.70	555.0
116.80	1170.0
117.70	619.0
118.70	902.0
123.60	103.0
127.70	530.0
128.70	237.0
129.80	577.0
130.80	225.0
134.80	466.0
136.90	433.0
140.00	118.0
140.80	2120.0
141.90	335.0
143.00	1950.0
143.90	137.0
144.90	163.0
145.80	250.0
146.80	173.0
148.00	347.0
148.80	197.0
149.90	101.0
152.90	103.0
155.00	401.0
156.70	232.0
160.80	269.0
169.10	125.0
170.70	260.0
172.10	486.0
172.70	917.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

173.90	118288.0
174.90	8639.0
175.90	116016.0
176.90	7594.0
178.00	120.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

m/z	Abundance
35.90	1284.0
36.90	7626.0
37.90	7575.0
38.90	3051.0
39.90	554.0
41.00	351.0
41.90	251.0
43.10	190.0
44.00	2241.0
44.90	1551.0
45.80	360.0
47.00	1462.0
47.80	1149.0
48.90	5975.0
49.90	29568.0
50.90	10030.0
51.90	307.0
54.80	350.0
55.90	2840.0
57.00	4827.0
57.90	357.0
59.90	1659.0
61.00	7979.0
61.90	8014.0
62.90	6336.0
64.00	733.0
64.90	121.0
67.00	702.0
67.90	15832.0
68.90	17368.0
69.90	1522.0
71.00	144.0
71.80	982.0
73.00	7229.0
74.00	29048.0
74.90	80408.0
76.00	6550.0
77.00	948.0
77.80	665.0
78.90	7235.0
80.00	1967.0
80.80	6927.0
81.90	1541.0
82.90	203.0
87.00	6059.0
87.90	5342.0
89.80	122.0
90.70	843.0
92.00	4919.0
92.90	6504.0
94.00	18200.0
95.00	148608.0
96.00	9565.0
96.90	260.0
103.80	949.0
105.00	162.0
105.80	874.0
106.60	203.0
109.70	125.0
110.80	409.0
111.80	148.0
112.80	233.0
114.70	254.0
115.80	755.0
116.90	1097.0
117.80	742.0
118.80	1009.0
127.80	670.0
128.90	164.0
129.90	522.0
130.60	146.0
131.00	141.0
134.90	466.0
136.90	475.0
140.80	1851.0
141.80	258.0
142.80	2136.0
145.60	202.0
148.00	321.0
148.80	108.0
151.80	199.0
153.00	142.0
154.90	355.0
156.80	277.0
160.70	145.0
169.30	138.0
171.00	202.0
171.90	403.0
173.00	988.0
173.90	108168.0
174.90	7680.0
175.90	102912.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

176.90	6773.0
177.80	124.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28