

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011025\
 Data File : VU062682.D
 Acq On : 10 Jan 2025 14:24
 Operator : MD/SY
 Sample : Q1058-12
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH8Y1

Quant Time: Jan 10 22:45:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 10 22:42:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	98192	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	88522	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	43393	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	22040	4.053	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.000%
7) Chloroethane-d5	1.904	69	20306	4.642	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.800%
11) 1,1-Dichloroethene-d2	2.554	65	12250	4.348	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.000%
20) 2-Butanone-d5	4.605	46	58418	52.559	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.120%
24) Chloroform-d	5.049	84	64878	4.849	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
26) 1,2-Dichloroethane-d4	5.692	65	37214	5.128	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.600%
32) Benzene-d6	5.714	84	109009	5.038	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.800%
36) 1,2-Dichloropropane-d6	6.679	67	31564	5.261	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%
41) Toluene-d8	7.888	98	103198	4.895	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
43) trans-1,3-Dichloroprop...	8.171	79	15617	5.222	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.400%
46) 2-Hexanone-d5	8.621	63	49757	52.900	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.800%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	27228	5.419	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.400%
66) 1,2-Dichlorobenzene-d4	12.183	152	38413	5.321	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.400%
Target Compounds					Qvalue	
34) Trichloroethene	6.534	95	3010	0.346	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011025\
Data File : VU062682.D
Acq On : 10 Jan 2025 14:24
Operator : MD/SY
Sample : Q1058-12
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH8Y1

Quant Time: Jan 10 22:45:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 10 22:42:03 2025
Response via : Initial Calibration

