

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011025\
 Data File : VU062696.D
 Acq On : 10 Jan 2025 20:58
 Operator : MD/SY
 Sample : VU0110WBL02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK137

Quant Time: Jan 11 00:39:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 10 22:42:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	97159	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	90996	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	42393	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	22721	4.222	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.400%
7) Chloroethane-d5	1.904	69	20442	4.722	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.400%
11) 1,1-Dichloroethene-d2	2.557	65	12182	4.370	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.400%
20) 2-Butanone-d5	4.605	46	61773	56.168	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.340%
24) Chloroform-d	5.049	84	64135	4.844	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.689	65	38702	5.389	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.800%
32) Benzene-d6	5.714	84	110900	4.986	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
36) 1,2-Dichloropropane-d6	6.679	67	32512	5.272	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.400%
41) Toluene-d8	7.888	98	103287	4.766	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.400%
43) trans-1,3-Dichloroprop...	8.171	79	15981	5.198	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.000%
46) 2-Hexanone-d5	8.621	63	51094	52.845	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.680%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	29258	5.665	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.200%
66) 1,2-Dichlorobenzene-d4	12.183	152	37632	5.336	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011025\
Data File : VU062696.D
Acq On : 10 Jan 2025 20:58
Operator : MD/SY
Sample : VU0110WBL02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK137

Quant Time: Jan 11 00:39:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 10 22:42:03 2025
Response via : Initial Calibration

