

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011322\
 Data File : VU046757.D
 Acq On : 13 Jan 2022 15:27
 Operator : SY/MD
 Sample : N1101-08DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFX0DL

Quant Time: Jan 13 23:13:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 10 00:25:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	73880	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	76898	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	39094	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	17514	3.010	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.200%
7) Chloroethane-d5	1.916	69	16387	3.576	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	71.600%
11) 1,1-Dichloroethene-d2	2.568	65	7900	3.219	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.400%
20) 2-Butanone-d5	4.639	46	87028	51.557	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.120%
24) Chloroform-d	5.067	84	45682	4.284	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
26) 1,2-Dichloroethane-d4	5.703	65	26725	4.384	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
32) Benzene-d6	5.729	84	83289	3.885	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.800%
36) 1,2-Dichloropropane-d6	6.694	67	27346	4.245	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.000%
41) Toluene-d8	7.899	98	68404	3.516	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.400%
43) trans-1,3-Dichloroprop...	8.179	79	11297	3.792	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.800%
46) 2-Hexanone-d5	8.636	63	74525	42.608	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.220%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	28946	4.378	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	29840	4.144	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.800%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.584	96	3820	0.588	ug/L	# 35
19) 1,1-Dichloroethane	3.877	63	20189	1.581	ug/L	97
29) 1,1,1-Trichloroethane	5.321	97	1862	0.146	ug/L	93
34) Trichloroethene	6.542	95	75982	9.434	ug/L	97
47) Tetrachloroethene	8.555	164	1616	0.284	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011322\
Data File : VU046757.D
Acq On : 13 Jan 2022 15:27
Operator : SY/MD
Sample : N1101-08DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFX0DL

Quant Time: Jan 13 23:13:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jan 10 00:25:50 2022
Response via : Initial Calibration

