

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011722\
 Data File : VU046766.D
 Acq On : 17 Jan 2022 15:24
 Operator : SY/MD
 Sample : N1168-01ME
 Misc : 5.76g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLW7ME

Quant Time: Jan 18 02:17:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jan 18 02:15:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	671590	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	643919	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	322687	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	186605	31.841	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.680%
7) Chloroethane-d5	1.916	69	127922	29.383	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	58.760%#
11) 1,1-Dichloroethene-d2	2.552	63	250672	25.149	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	50.300%#
21) 2-Butanone-d5	4.649	46	300224	69.995	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	70.000%
24) Chloroform-d	5.063	84	350464	33.923	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	67.840%#
26) 1,2-Dichloroethane-d4	5.703	65	239538	36.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	73.120%
32) Benzene-d6	5.726	84	740983	36.874	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	73.740%
36) 1,2-Dichloropropane-d6	6.690	67	224499	36.615	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	73.220%
41) Toluene-d8	7.899	98	696791	38.787	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.580%#
43) trans-1,3-Dichloroprop...	8.182	79	118798	40.217	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.440%
47) 2-Hexanone-d5	8.671	63	267539	82.135	ug/L	0.04
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.130%
56) 1,1,2,2-Tetrachloroeth...	10.767	84	341905	36.005	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	72.000%
66) 1,2-Dichlorobenzene-d4	12.198	152	271873	40.717	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.440%
Target Compounds						
					Qvalue	
46) Tetrachloroethene	8.552	164	10571	2.773	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011722\
Data File : VU046766.D
Acq On : 17 Jan 2022 15:24
Operator : SY/MD
Sample : N1168-01ME
Misc : 5.76g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLW7ME

Quant Time: Jan 18 02:17:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jan 18 02:15:56 2022
Response via : Initial Calibration

