

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012424\
 Data File : VU057619.D
 Acq On : 24 Jan 2024 15:21
 Operator : MD/SY
 Sample : P1228-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B98

Quant Time: Jan 25 01:54:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 25 01:52:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	57019	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	51403	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	23880	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	14749	4.695	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.800%
7) Chloroethane-d5	1.914	69	12801	4.863	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.200%
11) 1,1-Dichloroethene-d2	2.563	65	7097	4.804	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.000%
20) 2-Butanone-d5	4.640	46	34268	57.821	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.640%
24) Chloroform-d	5.055	84	31954	5.011	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.698	65	15410	5.318	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.400%
32) Benzene-d6	5.721	84	64189	5.324	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.400%
36) 1,2-Dichloropropane-d6	6.685	67	20029	5.507	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.200%
41) Toluene-d8	7.894	98	58328	5.096	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.000%
43) trans-1,3-Dichloroprop...	8.177	79	7476	5.158	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.200%
46) 2-Hexanone-d5	8.631	63	25873	55.947	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.900%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	13132	5.387	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	19353	5.775	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.400%
Target Compounds						Qvalue
47) Tetrachloroethene	8.547	164	3124	0.875	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012424\
Data File : VU057619.D
Acq On : 24 Jan 2024 15:21
Operator : MD/SY
Sample : P1228-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B98

Quant Time: Jan 25 01:54:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 25 01:52:13 2024
Response via : Initial Calibration

