

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012424\
 Data File : VU057635.D
 Acq On : 24 Jan 2024 22:10
 Operator : MD/SY
 Sample : P1227-08DL 200X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B84DL

Quant Time: Jan 25 04:58:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 25 04:41:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	56350	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	51456	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	23661	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	14318	4.612	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.200%
7) Chloroethane-d5	1.914	69	12102	4.652	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.000%
11) 1,1-Dichloroethene-d2	2.563	65	6633	4.544	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.800%
20) 2-Butanone-d5	4.647	46	33226	56.728	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.460%
24) Chloroform-d	5.055	84	31281	4.964	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.698	65	15525	5.421	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.400%
32) Benzene-d6	5.721	84	61783	5.119	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
36) 1,2-Dichloropropane-d6	6.685	67	18932	5.200	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.000%
41) Toluene-d8	7.894	98	56215	4.906	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
43) trans-1,3-Dichloroprop...	8.177	79	7050	4.859	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.200%
46) 2-Hexanone-d5	8.631	63	25649	55.405	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.820%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	12866	5.272	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	17966	5.411	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.200%
Target Compounds						
					Qvalue	
47) Tetrachloroethene	8.547	164	153525	42.938	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012424\
Data File : VU057635.D
Acq On : 24 Jan 2024 22:10
Operator : MD/SY
Sample : P1227-08DL 200X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B84DL

Quant Time: Jan 25 04:58:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 25 04:41:48 2024
Response via : Initial Calibration

