

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012424\
 Data File : VU057646.D
 Acq On : 25 Jan 2024 02:36
 Operator : MD/SY
 Sample : P1229-23 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BA8

Quant Time: Jan 25 05:00:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 25 04:41:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	52906	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	47150	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	21402	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	12834	4.403	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.000%	
7) Chloroethane-d5	1.914	69	11484	4.702	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.000%	
11) 1,1-Dichloroethene-d2	2.563	65	6165	4.498	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.000%	
20) 2-Butanone-d5	4.650	46	32973	59.961	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	119.920%	
24) Chloroform-d	5.055	84	28938	4.891	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.800%	
26) 1,2-Dichloroethane-d4	5.698	65	14714	5.472	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.400%	
32) Benzene-d6	5.721	84	57270	5.178	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.600%	
36) 1,2-Dichloropropane-d6	6.685	67	18280	5.480	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	109.600%	
41) Toluene-d8	7.894	98	50603	4.820	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.400%	
43) trans-1,3-Dichloroprop...	8.177	79	6470	4.866	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	97.400%	
46) 2-Hexanone-d5	8.634	63	24250	57.167	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	114.340%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	11930	5.335	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.800%	
66) 1,2-Dichlorobenzene-d4	12.190	152	16259	5.413	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	108.200%	
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	2248	0.545	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012424\
Data File : VU057646.D
Acq On : 25 Jan 2024 02:36
Operator : MD/SY
Sample : P1229-23 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BA8

Quant Time: Jan 25 05:00:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 25 04:41:48 2024
Response via : Initial Calibration

