

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012424\
 Data File : VU057647.D
 Acq On : 25 Jan 2024 03:00
 Operator : MD/SY
 Sample : P1229-22 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BA7

Quant Time: Jan 25 05:00:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 25 04:41:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	55738	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	49415	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	22736	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	12769	4.158	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	83.200%	
7) Chloroethane-d5	1.914	69	11277	4.382	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	87.600%	
11) 1,1-Dichloroethene-d2	2.560	65	6107	4.229	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	84.600%	
20) 2-Butanone-d5	4.650	46	32536	56.160	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	112.320%	
24) Chloroform-d	5.055	84	28767	4.615	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.200%	
26) 1,2-Dichloroethane-d4	5.698	65	14708	5.192	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.800%	
32) Benzene-d6	5.721	84	56419	4.867	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.400%	
36) 1,2-Dichloropropane-d6	6.682	67	17858	5.108	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.200%	
41) Toluene-d8	7.894	98	50665	4.605	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.000%	
43) trans-1,3-Dichloroprop...	8.177	79	6540	4.694	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	93.800%	
46) 2-Hexanone-d5	8.634	63	24834	55.861	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	111.720%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	12419	5.299	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.000%	
66) 1,2-Dichlorobenzene-d4	12.190	152	16502	5.172	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.400%	
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	2294	0.528	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012424\
Data File : VU057647.D
Acq On : 25 Jan 2024 03:00
Operator : MD/SY
Sample : P1229-22 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BA7

Quant Time: Jan 25 05:00:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 25 04:41:48 2024
Response via : Initial Calibration

