

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012524\
 Data File : VU057666.D
 Acq On : 25 Jan 2024 14:23
 Operator : MD/SY
 Sample : P1228-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BA1

Quant Time: Jan 25 20:08:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 25 04:41:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	68228	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	61114	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28497	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	13664	3.635	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.600%
7) Chloroethane-d5	1.911	69	12204	3.874	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.400%
11) 1,1-Dichloroethene-d2	2.563	65	6683	3.781	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.600%
20) 2-Butanone-d5	4.634	46	30903	43.577	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.160%
24) Chloroform-d	5.055	84	31509	4.129	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.600%
26) 1,2-Dichloroethane-d4	5.698	65	15211	4.387	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.800%
32) Benzene-d6	5.721	84	62316	4.347	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.000%
36) 1,2-Dichloropropane-d6	6.685	67	19329	4.470	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.400%
41) Toluene-d8	7.894	98	57101	4.196	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	8.177	79	6898	4.003	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.000%
46) 2-Hexanone-d5	8.631	63	23449	42.648	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.300%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	12708	4.385	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	18304	4.577	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012524\
Data File : VU057666.D
Acq On : 25 Jan 2024 14:23
Operator : MD/SY
Sample : P1228-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BA1

Quant Time: Jan 25 20:08:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 25 04:41:48 2024
Response via : Initial Calibration

