

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012524\
 Data File : VU057663.D
 Acq On : 25 Jan 2024 13:10
 Operator : MD/SY
 Sample : P1227-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B92

Quant Time: Jan 25 20:07:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 25 04:41:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	69089	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	61457	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28753	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	13904	3.653	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.000%
7) Chloroethane-d5	1.911	69	12518	3.925	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.400%
11) 1,1-Dichloroethene-d2	2.563	65	7009	3.916	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.400%
20) 2-Butanone-d5	4.631	46	34618	48.207	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.420%
24) Chloroform-d	5.055	84	32864	4.253	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
26) 1,2-Dichloroethane-d4	5.698	65	15815	4.504	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%
32) Benzene-d6	5.721	84	63755	4.423	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
36) 1,2-Dichloropropane-d6	6.685	67	19716	4.534	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.600%
41) Toluene-d8	7.894	98	58241	4.256	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	8.177	79	7280	4.201	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.000%
46) 2-Hexanone-d5	8.631	63	26422	47.787	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.580%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	13257	4.549	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	19007	4.710	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012524\
Data File : VU057663.D
Acq On : 25 Jan 2024 13:10
Operator : MD/SY
Sample : P1227-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B92

Quant Time: Jan 25 20:07:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 25 04:41:48 2024
Response via : Initial Calibration

