

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
 Data File : VU046954.D
 Acq On : 01 Feb 2022 10:44
 Operator : SY/MD
 Sample : VU0201WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK059

Quant Time: Feb 02 02:02:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 02 02:02:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	109000	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	124132	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	50730	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	54869	5.941	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.800%
7) Chloroethane-d5	1.922	69	42143	5.791	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.800%
11) 1,1-Dichloroethene-d2	2.575	65	20986	5.712	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	114.200%
20) 2-Butanone-d5	4.658	46	115046	42.045	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.100%
24) Chloroform-d	5.073	84	80806	4.984	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
26) 1,2-Dichloroethane-d4	5.710	65	45811	5.188	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%
32) Benzene-d6	5.735	84	161351	4.423	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
36) 1,2-Dichloropropane-d6	6.697	67	50899	4.647	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.902	98	126962	3.802	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.000%
43) trans-1,3-Dichloroprop...	8.185	79	18997	3.949	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.000%
46) 2-Hexanone-d5	8.642	63	114296	40.126	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.260%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	49460	4.497	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	42912	4.511	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
Data File : VU046954.D
Acq On : 01 Feb 2022 10:44
Operator : SY/MD
Sample : VU0201WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK059

Quant Time: Feb 02 02:02:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 02 02:02:08 2022
Response via : Initial Calibration

