

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
 Data File : VU046960.D
 Acq On : 01 Feb 2022 13:33
 Operator : SY/MD
 Sample : N1328-02DL 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HM1DL

Quant Time: Feb 02 02:04:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 02 02:02:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	105168	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	117601	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	49098	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	48765	5.472	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.400%
7) Chloroethane-d5	1.919	69	37858	5.392	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.800%
11) 1,1-Dichloroethene-d2	2.575	65	19003	5.361	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	107.200%
20) 2-Butanone-d5	4.652	46	107496	40.718	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.440%
24) Chloroform-d	5.070	84	76439	4.886	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.800%
26) 1,2-Dichloroethane-d4	5.706	65	41906	4.919	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
32) Benzene-d6	5.732	84	147329	4.263	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.200%
36) 1,2-Dichloropropane-d6	6.694	67	48203	4.645	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.800%
41) Toluene-d8	7.902	98	115130	3.639	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.800%
43) trans-1,3-Dichloroprop...	8.182	79	17246	3.784	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.600%
46) 2-Hexanone-d5	8.639	63	107152	39.707	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.420%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	44025	4.226	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	40346	4.382	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.600%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.588	96	30887	5.260	ug/L	95
19) 1,1-Dichloroethane	3.883	63	2707	0.254	ug/L	# 87
22) cis-1,2-Dichloroethene	4.674	96	52031	7.755	ug/L	96
29) 1,1,1-Trichloroethane	5.321	97	22233	2.099	ug/L	99
34) Trichloroethene	6.546	95	85946	11.956	ug/L	97
47) Tetrachloroethene	8.555	164	248357	42.478	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
Data File : VU046960.D
Acq On : 01 Feb 2022 13:33
Operator : SY/MD
Sample : N1328-02DL 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HM1DL

Quant Time: Feb 02 02:04:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 02 02:02:08 2022
Response via : Initial Calibration

