

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
 Data File : VU046976.D
 Acq On : 01 Feb 2022 21:01
 Operator : SY/MD
 Sample : VU0201WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK060

Quant Time: Feb 02 03:49:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 02 03:49:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	98026	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	116632	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	48590	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	46090	5.549	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.000%
7) Chloroethane-d5	1.919	69	36574	5.588	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.800%
11) 1,1-Dichloroethene-d2	2.575	65	17914	5.422	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	108.400%
20) 2-Butanone-d5	4.645	46	92902	37.754	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.500%
24) Chloroform-d	5.070	84	72479	4.970	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
26) 1,2-Dichloroethane-d4	5.707	65	40306	5.076	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
32) Benzene-d6	5.732	84	140329	4.094	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.800%
36) 1,2-Dichloropropane-d6	6.694	67	45361	4.407	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.200%
41) Toluene-d8	7.903	98	105510	3.363	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.200%#
43) trans-1,3-Dichloroprop...	8.182	79	16299	3.606	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.200%
46) 2-Hexanone-d5	8.639	63	86739	32.410	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	64.820%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	41620	4.028	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	37948	4.164	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
Data File : VU046976.D
Acq On : 01 Feb 2022 21:01
Operator : SY/MD
Sample : VU0201WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK060

Quant Time: Feb 02 03:49:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 02 03:49:05 2022
Response via : Initial Calibration

