

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
 Data File : VU046956.D
 Acq On : 01 Feb 2022 11:55
 Operator : SY/MD
 Sample : VIBLK029
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK029

Quant Time: Feb 02 02:03:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 02 02:02:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	116418	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	128829	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	50471	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	57356	5.814	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 116.200%		
7) Chloroethane-d5	1.909	69	44589	5.737	ug/L	-0.01
Spiked Amount	5.000	Range 65 - 130	Recovery	= 114.800%		
11) 1,1-Dichloroethene-d2	2.562	65	22569	5.752	ug/L	-0.02
Spiked Amount	5.000	Range 60 - 125	Recovery	= 115.000%		
20) 2-Butanone-d5	4.636	46	97468	33.352	ug/L	-0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	= 66.700%		
24) Chloroform-d	5.060	84	86507	4.995	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.000%		
26) 1,2-Dichloroethane-d4	5.703	65	48566	5.150	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 103.000%		
32) Benzene-d6	5.726	84	170428	4.501	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 90.000%		
36) 1,2-Dichloropropane-d6	6.694	67	54957	4.834	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 96.600%		
41) Toluene-d8	7.899	98	130922	3.778	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 75.600%		
43) trans-1,3-Dichloroprop...	8.179	79	19326	3.871	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 77.400%		
46) 2-Hexanone-d5	8.639	63	107690	36.429	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 72.860%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	48596	4.258	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 85.200%		
66) 1,2-Dichlorobenzene-d4	12.195	152	42644	4.505	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 90.200%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
Data File : VU046956.D
Acq On : 01 Feb 2022 11:55
Operator : SY/MD
Sample : VIBLK029
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK029

Quant Time: Feb 02 02:03:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 02 02:02:08 2022
Response via : Initial Calibration

