

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
 Data File : VU046985.D
 Acq On : 02 Feb 2022 00:36
 Operator : SY/MD
 Sample : N1349-11 100X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HQ0

Quant Time: Feb 02 04:08:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 02 03:49:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	96015	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	112983	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	46689	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	47054	5.783	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.600%
7) Chloroethane-d5	1.919	69	37057	5.781	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.600%
11) 1,1-Dichloroethene-d2	2.575	65	18324	5.662	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	113.200%
20) 2-Butanone-d5	4.645	46	114941	47.688	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.380%
24) Chloroform-d	5.067	84	73767	5.165	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.200%
26) 1,2-Dichloroethane-d4	5.706	65	40717	5.235	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.600%
32) Benzene-d6	5.732	84	144289	4.345	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.000%
36) 1,2-Dichloropropane-d6	6.694	67	47399	4.754	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.903	98	111185	3.658	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.200%
43) trans-1,3-Dichloroprop...	8.179	79	16744	3.824	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.400%
46) 2-Hexanone-d5	8.636	63	89452	34.503	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.000%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	42526	4.248	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	39147	4.471	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.400%
Target Compounds						Qvalue
34) Trichloroethene	6.542	95	1717	0.249	ug/L	87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020122\
Data File : VU046985.D
Acq On : 02 Feb 2022 00:36
Operator : SY/MD
Sample : N1349-11 100X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HQ0


Quant Time: Feb 02 04:08:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 02 03:49:05 2022
Response via : Initial Calibration





#34
 Trichloroethene
 Concen: 0.249 ug/L
 RT: 6.542 min Scan# 1
 Delta R.T. -0.003 min
 Lab File: VU046985.D
 Acq: 02 Feb 2022 00:36

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HQ0



Tgt Ion: 95 Resp: 1717

Ion	Ratio	Lower	Upper
95	100		
97	72.2	46.5	86.5
132	75.9	65.1	120.9
130	83.7	67.3	124.9

