

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020422\
 Data File : VU047048.D
 Acq On : 04 Feb 2022 15:57
 Operator : SY/MD
 Sample : N1348-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Feb 05 01:34:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 05 01:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	85510	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	101847	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	44318	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	59213	5.398	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 108.000%		
7) Chloroethane-d5	1.916	69	45214	5.339	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 106.800%		
11) 1,1-Dichloroethene-d2	2.572	65	22160	5.644	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 112.800%		
20) 2-Butanone-d5	4.629	46	122503	55.851	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 111.700%		
24) Chloroform-d	5.067	84	97736	5.586	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 111.800%		
26) 1,2-Dichloroethane-d4	5.706	65	49716	5.562	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 111.200%		
32) Benzene-d6	5.729	84	193497	5.425	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 108.400%		
36) 1,2-Dichloropropane-d6	6.694	67	59715	5.366	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 107.400%		
41) Toluene-d8	7.899	98	156263	4.939	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 98.800%		
43) trans-1,3-Dichloroprop...	8.179	79	22287	5.256	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 105.200%		
46) 2-Hexanone-d5	8.632	63	102743	52.313	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 104.620%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	49674	5.229	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 104.600%		
66) 1,2-Dichlorobenzene-d4	12.195	152	56696	6.283	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 125.600%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020422\
Data File : VU047048.D
Acq On : 04 Feb 2022 15:57
Operator : SY/MD
Sample : N1348-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Feb 05 01:34:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 05 01:18:17 2022
Response via : Initial Calibration

