

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020422\
 Data File : VU047051.D
 Acq On : 04 Feb 2022 17:10
 Operator : SY/MD
 Sample : N1349-15DL 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HQ4DL

Quant Time: Feb 05 01:35:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 05 01:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	94941	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	112674	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	47653	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	60522	4.969	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.400%
7) Chloroethane-d5	1.919	69	46787	4.976	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.600%
11) 1,1-Dichloroethene-d2	2.571	65	22585	5.181	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.600%
20) 2-Butanone-d5	4.639	46	128177	52.633	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.260%
24) Chloroform-d	5.067	84	97431	5.016	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
26) 1,2-Dichloroethane-d4	5.706	65	50171	5.056	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.732	84	193019	4.891	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.800%
36) 1,2-Dichloropropane-d6	6.693	67	59555	4.838	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.800%
41) Toluene-d8	7.902	98	155955	4.455	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
43) trans-1,3-Dichloroprop...	8.179	79	21595	4.604	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.000%
46) 2-Hexanone-d5	8.636	63	107639	49.539	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.080%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	50106	4.768	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	57183	5.894	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.800%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.584	96	7990	1.134	ug/L	# 1
22) cis-1,2-Dichloroethene	4.674	96	7068	0.914	ug/L	98
34) Trichloroethene	6.546	95	19148	2.393	ug/L	97
47) Tetrachloroethene	8.555	164	54213	8.732	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020422\
Data File : VU047051.D
Acq On : 04 Feb 2022 17:10
Operator : SY/MD
Sample : N1349-15DL 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HQ4DL

Quant Time: Feb 05 01:35:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 05 01:18:17 2022
Response via : Initial Calibration

