

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020422\
 Data File : VU047056.D
 Acq On : 04 Feb 2022 19:35
 Operator : SY/MD
 Sample : VIBLK039
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK039

Quant Time: Feb 05 01:36:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 05 01:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	89923	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	107980	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	46708	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	60386	5.235	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.600%
7) Chloroethane-d5	1.919	69	47323	5.314	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.200%
11) 1,1-Dichloroethene-d2	2.571	65	22558	5.463	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	109.200%
20) 2-Butanone-d5	4.636	46	128981	55.919	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.840%
24) Chloroform-d	5.067	84	97781	5.315	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.200%
26) 1,2-Dichloroethane-d4	5.706	65	51030	5.429	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.600%
32) Benzene-d6	5.732	84	194132	5.133	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.600%
36) 1,2-Dichloropropane-d6	6.693	67	59659	5.057	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.200%
41) Toluene-d8	7.902	98	157663	4.700	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	8.182	79	21642	4.814	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.200%
46) 2-Hexanone-d5	8.636	63	106530	51.160	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.320%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	51279	5.091	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	57222	6.017	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.400%#
Target Compounds						
					Qvalue	
47) Tetrachloroethene	8.555	164	9612	1.616	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020422\
Data File : VU047056.D
Acq On : 04 Feb 2022 19:35
Operator : SY/MD
Sample : VIBLK039
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK039

Quant Time: Feb 05 01:36:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 05 01:18:17 2022
Response via : Initial Calibration

