

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020422\
 Data File : VU047046.D
 Acq On : 04 Feb 2022 15:09
 Operator : SY/MD
 Sample : VU0204WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK063

Quant Time: Feb 05 01:34:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 05 01:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	96892	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	112107	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	50724	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	65300	5.254	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.000%
7) Chloroethane-d5	1.919	69	49990	5.210	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.200%
11) 1,1-Dichloroethene-d2	2.571	65	24247	5.450	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	109.000%
20) 2-Butanone-d5	4.636	46	135616	54.566	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.140%
24) Chloroform-d	5.067	84	102058	5.148	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.000%
26) 1,2-Dichloroethane-d4	5.703	65	54461	5.377	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
32) Benzene-d6	5.732	84	209829	5.344	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.800%
36) 1,2-Dichloropropane-d6	6.693	67	64390	5.257	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%
41) Toluene-d8	7.899	98	170599	4.898	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
43) trans-1,3-Dichloroprop...	8.182	79	23993	5.141	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.800%
46) 2-Hexanone-d5	8.636	63	113587	52.541	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.080%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	53300	5.097	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	61847	5.989	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020422\
Data File : VU047046.D
Acq On : 04 Feb 2022 15:09
Operator : SY/MD
Sample : VU0204WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK063

Quant Time: Feb 05 01:34:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 05 01:18:17 2022
Response via : Initial Calibration

