

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020425\
 Data File : VU063138.D
 Acq On : 04 Feb 2025 23:05
 Operator : MD/SY
 Sample : Q1227-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2988

Quant Time: Feb 05 08:06:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020425WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 05 07:57:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	83210	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	79551	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	33245	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	20492	4.305	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.200%	
7) Chloroethane-d5	1.901	69	19456	4.404	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	88.000%	
11) 1,1-Dichloroethene-d2	2.554	65	9375	4.412	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	88.200%	
20) 2-Butanone-d5	4.608	46	67408	43.021	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	86.040%	
24) Chloroform-d	5.049	84	44482	4.227	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.600%	
26) 1,2-Dichloroethane-d4	5.689	65	23501	4.342	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.800%	
32) Benzene-d6	5.714	84	88642	4.272	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.400%	
36) 1,2-Dichloropropane-d6	6.679	67	29745	4.348	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	87.000%	
41) Toluene-d8	7.888	98	74362	4.057	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.200%	
43) trans-1,3-Dichloroprop...	8.171	79	9814	3.829	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.600%	
46) 2-Hexanone-d5	8.621	63	50105	38.838	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	77.680%	
56) 1,1,2,2-Tetrachloroeth...	10.746	84	25008	4.142	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	82.800%	
66) 1,2-Dichlorobenzene-d4	12.184	152	25101	4.522	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	90.400%	
Target Compounds						
13) Acetone	2.615	43	13170	13.347	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020425\
Data File : VU063138.D
Acq On : 04 Feb 2025 23:05
Operator : MD/SY
Sample : Q1227-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2988

Quant Time: Feb 05 08:06:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020425WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 05 07:57:12 2025
Response via : Initial Calibration

