

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020525\
 Data File : VU063182.D
 Acq On : 06 Feb 2025 01:46
 Operator : MD/SY
 Sample : Q1226-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCZK4

Quant Time: Feb 06 08:11:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020425WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Feb 06 07:25:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	91738	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	86689	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	35213	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	23134	4.409	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.200%
7) Chloroethane-d5	1.901	69	23439	4.813	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.200%
11) 1,1-Dichloroethene-d2	2.557	65	10939	4.670	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.400%
20) 2-Butanone-d5	4.605	46	79225	45.862	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.720%
24) Chloroform-d	5.049	84	55401	4.775	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.600%
26) 1,2-Dichloroethane-d4	5.689	65	29210	4.895	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
32) Benzene-d6	5.714	84	108270	4.788	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.800%
36) 1,2-Dichloropropane-d6	6.679	67	36777	4.933	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.600%
41) Toluene-d8	7.888	98	89854	4.498	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%
43) trans-1,3-Dichloroprop...	8.171	79	11946	4.277	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.600%
46) 2-Hexanone-d5	8.621	63	58015	41.267	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.540%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	29948	4.551	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.000%
66) 1,2-Dichlorobenzene-d4	12.183	152	30116	5.122	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020525\
Data File : VU063182.D
Acq On : 06 Feb 2025 01:46
Operator : MD/SY
Sample : Q1226-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCZK4

Quant Time: Feb 06 08:11:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020425WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 06 07:25:52 2025
Response via : Initial Calibration

