

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020625\
 Data File : VU063206.D
 Acq On : 06 Feb 2025 19:30
 Operator : MD/SY
 Sample : Q1291-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0C04

Quant Time: Feb 07 07:33:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020425WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 07 07:28:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	77839	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	74745	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	28871	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18499	4.155	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.000%
7) Chloroethane-d5	1.901	69	19422	4.700	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.000%
11) 1,1-Dichloroethene-d2	2.554	65	8760	4.407	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.200%
20) 2-Butanone-d5	4.605	46	80521	54.935	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.880%
24) Chloroform-d	5.049	84	48800	4.957	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.689	65	26293	5.193	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%
32) Benzene-d6	5.714	84	92921	4.766	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
36) 1,2-Dichloropropane-d6	6.676	67	32037	4.984	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.600%
41) Toluene-d8	7.888	98	75511	4.384	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
43) trans-1,3-Dichloroprop...	8.171	79	10517	4.367	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.400%
46) 2-Hexanone-d5	8.621	63	57817	47.698	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.400%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	29230	5.152	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.000%
66) 1,2-Dichlorobenzene-d4	12.184	152	26389	5.474	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020625\
Data File : VU063206.D
Acq On : 06 Feb 2025 19:30
Operator : MD/SY
Sample : Q1291-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0C04

Quant Time: Feb 07 07:33:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020425WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 07 07:28:28 2025
Response via : Initial Calibration

