

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020625\
 Data File : VU063194.D
 Acq On : 06 Feb 2025 14:32
 Operator : MD/SY
 Sample : Q1226-13DL 10X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCZN3DL

Quant Time: Feb 07 07:30:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020425WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 07 07:28:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.235	114	78731	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.406	117	76219	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	31296	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	17963	3.989	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.800%
7) Chloroethane-d5	1.891	69	18738	4.483	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.600%
11) 1,1-Dichloroethene-d2	2.551	65	8726	4.340	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.800%
20) 2-Butanone-d5	4.605	46	70791	47.750	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.500%
24) Chloroform-d	5.049	84	47960	4.817	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
26) 1,2-Dichloroethane-d4	5.689	65	25358	4.952	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.000%
32) Benzene-d6	5.714	84	91554	4.605	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
36) 1,2-Dichloropropane-d6	6.676	67	31643	4.828	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.888	98	75332	4.289	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
43) trans-1,3-Dichloroprop...	8.171	79	10487	4.270	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.400%
46) 2-Hexanone-d5	8.621	63	53238	43.071	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.140%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26760	4.626	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.600%
66) 1,2-Dichlorobenzene-d4	12.184	152	26529	5.076	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.600%
Target Compounds						
					Qvalue	
22) cis-1,2-Dichloroethene	4.644	96	2376	0.389	ug/L	93
34) Trichloroethene	6.534	95	3376	0.530	ug/L	95
47) Tetrachloroethene	8.544	164	17634	3.941	ug/L	98
51) Chlorobenzene	9.441	112	3277	0.207	ug/L	98
67) 1,2-Dichlorobenzene	12.203	146	2415	0.257	ug/L	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020625\
Data File : VU063194.D
Acq On : 06 Feb 2025 14:32
Operator : MD/SY
Sample : Q1226-13DL 10X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCZN3DL

Quant Time: Feb 07 07:30:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR020425WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 07 07:28:28 2025
Response via : Initial Calibration









#22
cis-1,2-Dichloroethene
Concen: 0.389 ug/L
RT: 4.644 min Scan# 1043
Delta R.T. -0.006 min
Lab File: VU063194.D
Acq: 06 Feb 2025 14:32

Instrument :
MSVOA_U
ClientSampleId :
DCZN3DL

Tgt Ion: 96 Resp: 2376




Ion	Ratio	Lower	Upper
96	100		
61	121.6	91.8	170.6
98	61.0	44.4	82.4




#34
Trichloroethene
Concen: 0.530 ug/L
RT: 6.534 min Scan# 1631
Delta R.T. 0.003 min
Lab File: VU063194.D
Acq: 06 Feb 2025 14:32

Tgt Ion: 95 Resp: 3376

Ion	Ratio	Lower	Upper
95	100		
97	70.0	46.7	86.7
132	97.8	62.4	115.8
130	97.4	67.3	124.9









#47
Tetrachloroethene
Concen: 3.941 ug/L
RT: 8.544 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU063194.D
Acq: 06 Feb 2025 14:32




Instrument :
MSVOA_U
ClientSampleId :
DCZN3DL


Tgt Ion:164 Resp: 17634
Ion Ratio Lower Upper
164 100
129 99.9 68.6 127.4
131 94.2 66.5 123.5
166 130.9 89.9 167.0



#51
Chlorobenzene
Concen: 0.207 ug/L
RT: 9.441 min Scan# 2535
Delta R.T. 0.003 min
Lab File: VU063194.D
Acq: 06 Feb 2025 14:32


Tgt Ion:112 Resp: 3277
Ion Ratio Lower Upper
112 100
114 32.3 22.6 42.0
77 61.0 50.8 76.2





#67
 1,2-Dichlorobenzene
 Concen: 0.257 ug/L
 RT: 12.203 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VU063194.D
 Acq: 06 Feb 2025 14:32

Instrument :
 MSVOA_U
 ClientSampleId :
 DCZN3DL



Tgt Ion:146 Resp: 2415
 Ion Ratio Lower Upper
 146 100
 111 45.8 34.1 51.1
 148 54.2 52.5 78.7

