

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021125\
 Data File : VU063237.D
 Acq On : 11 Feb 2025 15:45
 Operator : MD/SY
 Sample : Q1172-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-RVOA-WS

Quant Time: Feb 12 03:19:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 11 08:42:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.103	96	54333	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	18048	1.007	ug/l	0.00
Spiked Amount 1.000			Recovery	=	101.000%	
68) 1,2-Dichlorobenzene-d4	12.184	152	18316	0.983	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.000%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.599	62	289938	14.415	ug/l	99
9) 1,1-Dichloroethene	2.570	96	150522	10.920	ug/l	97
15) Methylene Chloride	3.030	84	256157	15.042	ug/l	98
16) trans-1,2-Dichloroethene	3.341	96	90153	5.730	ug/l	99
22) cis-1,2-Dichloroethene	4.650	96	295505	17.384	ug/l	88
28) 1,1,1-Trichloroethane	5.303	97	189267	7.807	ug/l	100
30) Carbon Tetrachloride	5.512	117	326894	15.724	ug/l	97
35) Benzene	5.763	78	181072	2.712	ug/l	99
36) 1,2-Dichloroethane	5.785	62	47795	2.481	ug/l	98
37) Trichloroethene	6.531	130	257443	16.215	ug/l	100
38) 1,2-Dichloropropane	6.779	63	49067	2.808	ug/l	99
49) Toluene	7.959	92	523597	13.638	ug/l	98
52) 1,1,2-Trichloroethane	8.386	97	94081	7.886	ug/l	98
58) Tetrachloroethene	8.541	164	193632	14.799	ug/l	98
59) Chlorobenzene	9.434	112	303811	7.499	ug/l	99
63) Ethyl Benzene	9.560	91	352913	5.051	ug/l	100
64) m/p-Xylenes	9.682	106	65739	2.519	ug/l	98
65) o-Xylene	10.090	106	67433	2.639	ug/l	99
66) Styrene	10.103	104	309385	7.609	ug/l	94
83) 1,4-Dichlorobenzene	11.827	146	426032	13.875	ug/l	99
85) 1,2-Dichlorobenzene	12.200	146	409950	13.589	ug/l	99
87) 1,2,4-Trichlorobenzene	13.830	180	185597	12.614	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021125\
Data File : VU063237.D
Acq On : 11 Feb 2025 15:45
Operator : MD/SY
Sample : Q1172-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-RVOA-WS

Quant Time: Feb 12 03:19:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

