

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
 Data File : VU063254.D
 Acq On : 12 Feb 2025 14:58
 Operator : MD/SY
 Sample : Q1172-10DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WSDL

Quant Time: Feb 13 03:12:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 11 08:42:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.107	96	48656	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	12366	0.771	ug/l	0.00
Spiked Amount 1.000			Recovery	=	77.000%	
68) 1,2-Dichlorobenzene-d4	12.190	152	13603	0.815	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.570	101	49347	4.073	ug/l	99
24) tert-Butyl Alcohol	3.171	59	11052	3.957	ug/l #	90
25) Methyl tert-Butyl Ether	3.351	73	44947	1.458	ug/l	97
31) Isopropyl Ether	3.978	45	76959	2.030	ug/l	96
32) Ethyl-t-butyl ether	4.486	59	334596	9.705	ug/l	99
33) Tert-Amyl methyl ether	5.926	73	269177	8.936	ug/l	100
73) n-propylbenzene	10.897	120	29507	1.916	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063254.D
Acq On : 12 Feb 2025 14:58
Operator : MD/SY
Sample : Q1172-10DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Quant Time: Feb 13 03:12:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

