

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022522\
 Data File : VU047265.D
 Acq On : 25 Feb 2022 11:36
 Operator : SY/MD
 Sample : VU0225WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK071

Quant Time: Feb 28 02:01:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Feb 28 02:01:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	345921	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	342122	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	200246	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	141134	44.923	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.840%
7) Chloroethane-d5	1.919	69	112252	48.199	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.400%
11) 1,1-Dichloroethene-d2	2.572	63	194884	37.433	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.860%
21) 2-Butanone-d5	4.642	46	274116	112.502	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.500%
24) Chloroform-d	5.067	84	247090	48.447	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.900%
26) 1,2-Dichloroethane-d4	5.707	65	158962	48.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.120%
32) Benzene-d6	5.732	84	524808	46.141	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.280%
36) 1,2-Dichloropropane-d6	6.694	67	164166	44.507	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.020%
41) Toluene-d8	7.900	98	499050	46.082	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.160%
43) trans-1,3-Dichloroprop...	8.179	79	80275	41.529	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.060%
47) 2-Hexanone-d5	8.636	63	220882	104.527	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.530%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	254318	47.285	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.580%
66) 1,2-Dichlorobenzene-d4	12.195	152	207737	43.742	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022522\
Data File : VU047265.D
Acq On : 25 Feb 2022 11:36
Operator : SY/MD
Sample : VU0225WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK071

Quant Time: Feb 28 02:01:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Feb 28 02:01:00 2022
Response via : Initial Calibration

