

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022522\
 Data File : VU047269.D
 Acq On : 25 Feb 2022 13:46
 Operator : SY/MD
 Sample : N1631-15ME
 Misc : 5.88g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBRJ1ME

Quant Time: Feb 28 02:02:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Feb 28 02:01:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	226396	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	229632	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.822	152	129400	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	46337	22.536	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	45.080%	#	
7) Chloroethane-d5	1.916	69	25003	16.404	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	32.800%	#	
11) 1,1-Dichloroethene-d2	2.543	63	170885	50.152	ug/L	-0.03
Spiked Amount 50.000	Range 60 - 125		Recovery =	100.300%		
21) 2-Butanone-d5	4.655	46	178854	112.159	ug/L	0.02
Spiked Amount 100.000	Range 40 - 130		Recovery =	112.160%		
24) Chloroform-d	5.067	84	182970	54.815	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	109.620%		
26) 1,2-Dichloroethane-d4	5.703	65	122283	57.076	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	114.160%		
32) Benzene-d6	5.723	84	393625	51.560	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	103.120%		
36) 1,2-Dichloropropane-d6	6.694	67	123238	49.778	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	99.560%		
41) Toluene-d8	7.903	98	383617	52.776	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	105.560%		
43) trans-1,3-Dichloroprop...	8.186	79	60771	46.840	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	93.680%		
47) 2-Hexanone-d5	8.687	63	151091	106.526	ug/L	0.05
Spiked Amount 100.000	Range 45 - 130		Recovery =	106.530%		
56) 1,1,2,2-Tetrachloroeth...	10.774	84	184259	51.042	ug/L	0.02
Spiked Amount 50.000	Range 65 - 120		Recovery =	102.080%		
66) 1,2-Dichlorobenzene-d4	12.201	152	152204	49.595	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	99.200%		
Target Compounds						Qvalue
3) Chloromethane	1.520	50	4753	2.277	ug/L	97
6) Bromomethane	1.861	94	2191	1.791	ug/L	98
22) 2-Butanone	4.739	43	22553	14.470	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022522\
Data File : VU047269.D
Acq On : 25 Feb 2022 13:46
Operator : SY/MD
Sample : N1631-15ME
Misc : 5.88g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBRJ1ME

Quant Time: Feb 28 02:02:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Feb 28 02:01:00 2022
Response via : Initial Calibration

