

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022522\
 Data File : VU047264.D
 Acq On : 25 Feb 2022 11:09
 Operator : SY/MD
 Sample : VU0225MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 VBLK070

Quant Time: Feb 28 02:01:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Feb 28 02:01:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	345299	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	335914	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	199758	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	150385	47.954	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery =	95.900%		
7) Chloroethane-d5	1.919	69	119446	51.381	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery =	102.760%		
11) 1,1-Dichloroethene-d2	2.575	63	206036	39.646	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	79.300%		
21) 2-Butanone-d5	4.639	46	278757	114.613	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery =	114.610%		
24) Chloroform-d	5.067	84	261687	51.401	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	102.800%		
26) 1,2-Dichloroethane-d4	5.706	65	166068	50.821	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	101.640%		
32) Benzene-d6	5.732	84	553599	49.572	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	99.140%		
36) 1,2-Dichloropropane-d6	6.694	67	170642	47.118	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery =	94.240%		
41) Toluene-d8	7.899	98	526577	49.523	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	99.040%		
43) trans-1,3-Dichloroprop...	8.182	79	84981	44.776	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	89.560%		
47) 2-Hexanone-d5	8.636	63	224052	107.987	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery =	107.990%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	261922	49.599	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery =	99.200%		
66) 1,2-Dichlorobenzene-d4	12.195	152	221887	46.835	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	93.680%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022522\
 Data File : VU047264.D
 Acq On : 25 Feb 2022 11:09
 Operator : SY/MD
 Sample : VU0225MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK070

Quant Time: Feb 28 02:01:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Feb 28 02:01:00 2022
 Response via : Initial Calibration

