

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030222\
 Data File : VU047294.D
 Acq On : 02 Mar 2022 16:26
 Operator : SY/MD
 Sample : N1745-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0AA2

Quant Time: Mar 02 23:57:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 02 23:54:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	148226	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	151109	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	93690	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	51440	4.109	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.200%
7) Chloroethane-d5	1.919	69	41865	4.220	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.400%
11) 1,1-Dichloroethene-d2	2.572	65	21647	4.181	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.600%
20) 2-Butanone-d5	4.649	46	157831	52.747	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.500%
24) Chloroform-d	5.067	84	91617	4.406	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	5.707	65	51594	4.462	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
32) Benzene-d6	5.729	84	189053	4.478	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
36) 1,2-Dichloropropane-d6	6.694	67	59572	4.434	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.600%
41) Toluene-d8	7.899	98	175937	4.260	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	8.182	79	25634	4.209	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.200%
46) 2-Hexanone-d5	8.636	63	138570	43.213	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.420%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	56353	4.615	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	12.192	152	75742	4.506	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.200%
Target Compounds						Qvalue
3) Chloromethane	1.527	50	3841	0.283	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030222\
Data File : VU047294.D
Acq On : 02 Mar 2022 16:26
Operator : SY/MD
Sample : N1745-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0AA2

Quant Time: Mar 02 23:57:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 02 23:54:47 2022
Response via : Initial Calibration

