

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030222\
 Data File : VU047298.D
 Acq On : 02 Mar 2022 18:00
 Operator : SY/MD
 Sample : N1745-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0AA6

Quant Time: Mar 02 23:58:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 02 23:54:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	144781	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	149796	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	91519	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	47880	3.916	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.400%
7) Chloroethane-d5	1.919	69	40100	4.138	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.800%
11) 1,1-Dichloroethene-d2	2.572	65	20854	4.124	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.400%
20) 2-Butanone-d5	4.649	46	154222	52.768	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.540%
24) Chloroform-d	5.067	84	87825	4.324	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.400%
26) 1,2-Dichloroethane-d4	5.704	65	48643	4.307	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
32) Benzene-d6	5.729	84	181400	4.334	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	6.694	67	57567	4.322	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.400%
41) Toluene-d8	7.900	98	169686	4.144	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	8.179	79	23643	3.916	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.400%
46) 2-Hexanone-d5	8.636	63	131018	41.216	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.440%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	54443	4.498	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	71608	4.361	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.200%
Target Compounds						Qvalue
3) Chloromethane	1.524	50	5445	0.410	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030222\
Data File : VU047298.D
Acq On : 02 Mar 2022 18:00
Operator : SY/MD
Sample : N1745-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0AA6

Quant Time: Mar 02 23:58:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 02 23:54:47 2022
Response via : Initial Calibration

