

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030322\
 Data File : VU047315.D
 Acq On : 03 Mar 2022 15:12
 Operator : SY/MD
 Sample : VIBLK007
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK007

Quant Time: Mar 04 01:55:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 04 01:50:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	136863	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	140426	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	86384	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	43854	3.794	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.800%
7) Chloroethane-d5	1.919	69	36714	4.008	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.200%
11) 1,1-Dichloroethene-d2	2.572	65	19251	4.027	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.600%
20) 2-Butanone-d5	4.652	46	143512	51.944	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.880%
24) Chloroform-d	5.067	84	83167	4.331	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
26) 1,2-Dichloroethane-d4	5.707	65	45921	4.301	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.000%
32) Benzene-d6	5.732	84	168410	4.292	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.800%
36) 1,2-Dichloropropane-d6	6.694	67	53939	4.320	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.400%
41) Toluene-d8	7.899	98	153926	4.010	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.200%
43) trans-1,3-Dichloroprop...	8.182	79	22080	3.901	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.000%
46) 2-Hexanone-d5	8.636	63	121609	40.809	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.620%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	51757	4.561	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.200%
66) 1,2-Dichlorobenzene-d4	12.192	152	68326	4.408	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030322\
Data File : VU047315.D
Acq On : 03 Mar 2022 15:12
Operator : SY/MD
Sample : VIBLK007
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK007

Quant Time: Mar 04 01:55:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 04 01:50:04 2022
Response via : Initial Calibration

