

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030422\
 Data File : VU047339.D
 Acq On : 04 Mar 2022 19:05
 Operator : SY/MD
 Sample : N1810-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBD89

Quant Time: Mar 04 23:24:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 04 23:20:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	185668	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	176477	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	95794	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	64199	4.872	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	97.400%	
7) Chloroethane-d5	1.919	69	57594	5.239	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.800%	
11) 1,1-Dichloroethene-d2	2.571	65	27459	4.934	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	98.600%	
20) 2-Butanone-d5	4.639	46	188276	43.520	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	87.040%	
24) Chloroform-d	5.067	84	125642	5.160	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.200%	
26) 1,2-Dichloroethane-d4	5.706	65	70121	5.132	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.600%	
32) Benzene-d6	5.729	84	256841	5.167	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.400%	
36) 1,2-Dichloropropane-d6	6.693	67	82065	5.301	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	106.000%	
41) Toluene-d8	7.899	98	234925	5.006	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.200%	
43) trans-1,3-Dichloroprop...	8.182	79	33667	4.949	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	99.000%	
46) 2-Hexanone-d5	8.636	63	192895	49.824	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.640%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	77660	4.924	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.400%	
66) 1,2-Dichlorobenzene-d4	12.192	152	101516	5.408	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	108.200%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.607	62	1174	0.095	ug/L #	14
12) 1,1-Dichloroethene	2.584	96	3447	0.379	ug/L #	1
22) cis-1,2-Dichloroethene	4.671	96	26040	2.420	ug/L	91
34) Trichloroethene	6.546	95	401837	38.387	ug/L	93
47) Tetrachloroethene	8.555	164	2227	0.250	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030422\
Data File : VU047339.D
Acq On : 04 Mar 2022 19:05
Operator : SY/MD
Sample : N1810-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBD89

Quant Time: Mar 04 23:24:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 04 23:20:39 2022
Response via : Initial Calibration

