

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030422\
 Data File : VU047351.D
 Acq On : 04 Mar 2022 23:48
 Operator : SY/MD
 Sample : N1810-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBDA2

Quant Time: Mar 05 01:28:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 04 23:20:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	183632	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	174239	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	92694	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	57232	4.391	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.800%
7) Chloroethane-d5	1.919	69	52832	4.859	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.200%
11) 1,1-Dichloroethene-d2	2.572	65	24569	4.464	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.200%
20) 2-Butanone-d5	4.639	46	173446	40.536	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.080%
24) Chloroform-d	5.067	84	117247	4.869	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%
26) 1,2-Dichloroethane-d4	5.707	65	66406	4.914	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	5.732	84	237546	4.841	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	6.694	67	75697	4.953	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.000%
41) Toluene-d8	7.899	98	217413	4.692	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.800%
43) trans-1,3-Dichloroprop...	8.179	79	30490	4.540	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	8.636	63	184437	48.251	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.500%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	72656	4.666	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.400%
66) 1,2-Dichlorobenzene-d4	12.192	152	94066	5.179	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030422\
Data File : VU047351.D
Acq On : 04 Mar 2022 23:48
Operator : SY/MD
Sample : N1810-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBDA2

Quant Time: Mar 05 01:28:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR030422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 04 23:20:39 2022
Response via : Initial Calibration

